## **UT Southwestern Department of Radiology**

Protocol Name: Ankle/Foot CT With IV Contrast

Orderable Name: CT LOWER EXTREMITY LEFT W IV CONTRAST

**Epic Button: Ankle/Foot + IV** 

CT LOWER EXTREMITY RIGHT W IV CONTRAST

CTDIvol < 50 mGy

Adult Only

Indications: Suspected infection, inflammation, tumor

# Acquisitions: 1 **Active Protocol** 

| Oral Contrast: None | IV Contrast: Link to Contrast Information   | Other Contrast: None | Airway   |
|---------------------|---|----------------------|--|
|                     | Rate (ml/sec): 3  |                      |  |
|                     | Volume (ml): 60 - 75  |                      |  |
|                     | IV Access:  |                      | Other Notes  *Place a marker at the site of most concern.  Dual energy/Spectral scanner required. Photon counting scanner preferred unless gout is indicated.  Use Right/Left orderable based on protocol or side indicated in reason for exam.  Metal (FOV): Use 140 KVp. |
|                     | Power injection: 20g or larger strongly prefered (if 22g use reduce rate to 2.5 mL/sec) |                      |  |
|                     | Notes: Adjust contrast volume based on patient size.                                    |                      |  |
|                     |   |                      |  |
|                     |   |                      |  |
|                     |   |                      |  |

| Last Change: 1/13           | /2023 Last Review: 1/29/2025 Link   | ss: <u>General Statements</u>  |  |
|-----------------------------|---|--|--|
| Special<br>Instructions     | Use 5mm cor/sag if large patient or metal in FOV                            | . Do not repeat CT scan, recon soft tissue from 1st acquisition, send soft tissue kernal volume to TeraRecon Use 5mm cor/sag if large patient or metal in FOV. |  |
| Acq # / Series Name         | 1 60 Sec Delayed  | N/A 60 Sec Delayed   |  |
| Phase Timing                | 60 seconds  | N/A  |  |
| <b>Acquisition Protocol</b> |   | Recon Only   |  |
| Coverage                    | Distal tibia thru toes  | Same   |  |
| FOV                         | Whole ankle and foot  | Same   |  |
| Algorithm                   | Bone  | Soft Tissue  |  |
| <b>Axial Recons</b>         | 3 mm  | 3 mm, volume   |  |
| <b>Other Planar Recons</b>  | 3 mm coronal and sagittal   | 4 mm coronal and sagittal  |  |
| MIP Recons                  |   |  |  |
| †DECT Philips               | VNC, Gout maps (cor/sag), BM edema, mono-E<br>100, mono-E 140, SBI          |  |  |
| †DECT Siemens               | VNC, Gout maps (cor/sag), BM edema, mono-E<br>100, mono-E 140, low/high kVp |  |  |
| <b>†PC-CT Siemens</b>       |   |  |  |

