

Anatomy: **PELVIS**  
 Sub-Anatomy: **Appendicitis/Acute Abdomen**

- Exams  
 - Routine Coil: **SENSE XL Torso**

SEQUENCE - BASICS				VARIABLE PARAMETERS														
PLANE	SEQ	Slice / Gap (mm)	Misc / Comment	MTX	FOV (cm)	% RFOV	Scan %	Pixel size (mm)	TR	TE	FA/TI	Nex Avg Acq	NS	ETL Turbo Factor	Phase Encode	Scan TIME (min)	Pixel Shift BW-kHz	
<b>ROUTINE</b>																		
SAG	2D T2 SSFSE	5/1																
COR	2D T2 SSFSE	5/1																
AX	2D T2 SSFSE	5/1																
AX	2D T2 FS FSE	5/1																
AX	2D TOF	4/-2	No SAT bands															
AX	2D DWI FS	7/1	b=0,50,500,1000															
AX	2D T1 IP/OP	5/1																
AX	3D bFFE	2	Cover bladder only															
<b>↓ OPTIONAL ↓</b>																		

**ORDERABLE: MR ABDOMEN WO CONTRAST & MR PELVIS WO CONTRAST**

COR T2 planning

Axial planning

Sag T2 planning  
(cover past midline)

