

Anatomy: **Breast BIOPSY:** - Exams
 Sub-Anatomy: **Breast** - Routine

Coil: **7 Channel Breast Coil**

SEQUENCE - BASICS				VARIABLE PARAMETERS													
PLANE	SEQ	Slice / Gap (mm)	Misc. / Comment	MTX	FOV (cm)	% RFO V	Scan %	Pixel size (mm)	TR	TE	FA/TI	Nex Avg Acq	NS	ETL Turbo Factor	Phase Encode	Scan TIME (min)	Pixel Shift BW-kHz
ROUTINE																	
SAG	T1 PRE INJ	32	FOV PATIENT DEPENDENT						32	2.5							
SAG	T1 POST INJ	32	FOV PATIENT DEPENDENT						32	2.5							
AX	T1 POST INJ	30	FOV PATIENT DEPENDENT						64	2.5							

TIPS: