

# UT Southwestern Department of Radiology

**Protocol Name:** CT Orbits Without and With

**Orderable Name:** CT ORBITS W AND WO IV CONTRAST

Adult Only

**Epic Button:** Orbits W/O & W/

CTDIvol < 80 mGy

**Indications:** Grave's disease, vision loss, blurred vision, eye pain, malignancy, infection, trauma.

**# Acquisitions:** 2

Active Protocol

<b>Oral Contrast:</b> None	<b>IV Contrast:</b> <a href="#">Link to Contrast Information</a> <b>Rate (ml/sec):</b> 3 <b>Volume (ml):</b> 70 <b>IV Access:</b> Power injection: 20g or larger strongly preferred (if 22g use reduce rate to 2.5 mL/sec)  <b>Notes:</b> Saline Test Injection: 10 sec @ injection rate	<b>Other Contrast:</b> None	<b>Airway</b>
			<b>Other Notes</b> No habitus specific kVp adjustment.

Last Change: 1/25/2023

Last Review: 1/8/2025

[Links:](#)

<b>Special Instructions</b>		Do not repeat CT scan, recon soft tissue from 1st acquisition	
<b>Acq # / Series Name</b>	<b>1</b> Noncontrast	<b>N/A</b> Noncontrast	<b>2</b> Contrast
<b>Phase Timing</b>		N/A	65 seconds
<b>Acquisition Protocol</b>		<a href="#">Recon Only</a>	
<b>Coverage</b>	Frontal sinuses thru maxillary sinuses	Same	Frontal sinuses thru maxillary sinuses
<b>FOV</b>	All of sinus space (front of face)	Same	All of sinus space (front of face)
<b>Algorithm</b>	Bone	Soft Tissue	Soft Tissue
<b>Axial Recons</b>	2 mm	2 mm, volume	2 mm, volume
<b>Other Planar Recons</b>	2 mm coronal and sagittal	2 mm coronal and sagittal	2 mm coronal and sagittal
<b>MIP Recons</b>			
<b>†DECT Philips</b>			
<b>†DECT Siemens</b>			
<b>†PC-CT Siemens</b>			

† When dual energy (DE) or photon counting (PC) CT is used

