

Routine Breast: Screening/Diagnostic/

Anatomy: Saline Implants
 Sub-Anatomy: Breast

- Exams

- Routine

Coil: 7 Channel Breast Coil

SEQUENCE - BASICS				VARIABLE PARAMETERS													
PLANE	SEQ	Slice / Gap (mm)	Misc. / Comment	MTX	FOV (cm)	% RFO V	Scan %	Pixel size (mm)	TR	TE	FA/TI	Nex Avg Acq	NS	ETL Turbo Factor	Phase Encode	Scan TIME (min)	Pixel Shift BW-kHz
ROUTINE																	
AX	T1 VISTA 3D	200	FOV PATIENT DEPEND						2000	204							
AX	T1W 3D	200	FOV PATIENT DEPEND						400	15							
AX	Dyn-e Thrive 3D	200	FOV PATIENT DEPEND						7.3	3.6							

TIPS: