

Anatomy: **Orbits/Face/Neck**

- Exams

Sub-Anatomy: **Skull Base (19SBC)**

- Routine

Coil: **NV Head Matrix**

SEQUENCE - BASICS				VARIABLE PARAMETERS													
PLANE	SEQ	Slice / Gap (mm)	Misc / Comment	MTX	FOV (cm)	% RFOV	Scan %	Pixel size (mm)	TR	TE	FA/TI	Nex Avg Acq	NS	ETL Turbo Factor	Phase Encode	Scan TIME (min)	Pixel Shift BW-kHz
ROUTINE																	
AX	DWI	5/0.5	Whole Head	128x128	22												
AX	T1	3/0.3			16												
COR	T1	3/0.3			16												
AX	T2 FS	3/0.3	(Use STIR if metal present)		16												
COR	T2 FS/Dixon	3/0.3	(Use STIR if metal present) If DIXON send only Water/IP images		16												
AX	T2 FLAIR	5/0.5	Whole Head		22												
AX	T1 POST	3/0.3			16												
COR	T1 FS/DIXON POST	3/0.3			16												
SAG	T1 FS/DIXON POST	3/0.3	Whole Head		22												
AX	T1 FS/ DIXON POST	5/0.5	Whole Head		22												
↓ OPTIONAL ↓																	

ORDERABLE: MR NECK W AND WO CONTRAST.

TIPS:

- Cover from under the corpus callosum genu to just below the hard palate.
- Center on lesion, do not image air.
- If metal is present, use STIR instead of T2 FS.